CALIFORNIA AIR RESOURCES BOARD
MONTHLY QUALITY CONTROL MAINTENANCE CHECK SHEET
ENVIRONICS 9100 GAS CALIBRATION SYSTEM

Location: _____ Month/Year: _____
Station Number: _____ Technician: _____
Property Number: _____ Agency: _____

OPERATOR INSTRUCTIONS:

1. Daily checks: Check power.
Check 9100 display. Verify instrument is in TIMER CONROL MODE.
Check 9100 time and date.
Review calibration data.

2. Weekly Checks: Record zero and precision/span readings for all instruments on appropriate maintenance form.

3. Monthly Checks: Check and record superblend output pressure _____
Check and record superblend cylinder pressure _____
Check and record superblend certification date _____
Verify that the cooling fan is operating.
Inspect air and gas line connections. Replace or repair as necessary.

4. Semi-Annual: Perform 9100 calibration/verification. Date last cal: _____

5. Annual Checks: Change particulate filter on Port 1.

Date	Comments or Maintenance Performed:

Reviewed by: _____ Date: _____