

**CALIFORNIA AIR RESOURCES BOARD
MONTHLY QUALITY CONTROL MAINTENANCE CHECK SHEET
Met-One BAM-1020 AIR SAMPLER**

Location: _____ Month/Year: _____
 Station Number: _____ Technician: _____
 Property Number: _____ Agency: _____

Instrument Checks:

- 1) Daily checks: Review station data logger values for correct operation of BAM-1020.
- 2) Weekly checks: Check filter tape & replace when necessary (approx. 2 months per roll).
- 3) Bi-weekly checks: Perform BAM-1020 flow and leak check.
- 4) Monthly checks: Complete and submit this Monthly Quality Control Check Sheet.
 Thoroughly clean both PM2.5 SCC and PM10 FRM inlets.
 Check pump muffler and replace when needed.
 Download and submit data from BAM-1020 data logger.
- 5) Semi-annual: Calibrate BAM-1020 Date last performed: _____
 Value of Zero Background: _____ Date last performed: _____
 Record data logger settings for max and min values: _____

Bi-Weekly Sampler Flow Rate, Ambient Temp and Pressure Check Results:

Date:	Standard	Sampler	Difference	Control Limits*
Flow Rate				16.00 to 17.34 L/min
Ambient Temp.				±2 °C
Ambient Press.				±10 mm Hg
Leak Check				< 1.0 L/min

Date:	Standard	Sampler	Difference	Control Limits*
Flow Rate				16.00 to 17.34 L/min
Ambient Temp.				±2 °C
Ambient Press.				±10 mm Hg
Leak Check				< 1.0 L/min

Date:	Standard	Sampler	Difference	Control Limits*
Flow Rate				16.00 to 17.34 L/min
Ambient Temp.				±2 °C
Ambient Press.				±10 mm Hg
Leak Check				< 1.0 L/min

*If check exceeds limits, investigate to determine cause and repeat check immediately.
 If second check also exceeds limits, request a re-calibration.

Transfer Standard Used

Make/Model	Serial/I.D.Number	Date Certified

Operator Comments: _____

Reviewed by: _____ Date: _____